



California Medical Device Recall Information



Recall Name

**Covidien Recalls the Pipeline Embolization and Alligator Retrieval Devices
Due to Risk of Coating Delamination and Detachment**

Recall Date	Product Description	Recalling Firm	Recall Reason
04/01/14	<ul style="list-style-type: none">Pipeline Embolization DeviceAlligator Retrieval Device	Covidien Irvine, CA	<i>Potential for the PTFE coating to delaminate and detach from the delivery wire.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	List of Affected Model and Lot Numbers	CA , nationwide	Distributed from: May 2013 to March 2014

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm392987.htm>